

Liquid Chromatography Mass Spectrometry

# LCMS Consumables



# LCMS-2020

Single Quadrupole Mass Spectrometer



## Inspection/Maintenance Schedule

	Part Number	Description	3 month Service	6 month Service	36 month Service	48 month Service	60 month Service
Vacuum System	220-97808-01	Pump Oil	X				
	220-91566-10	Oil Filter Element (MF30)		X			
	225-14179-01	Turbo Molecular Pump				X	
	225-09309-02	Rotary Pump, E2M28					X
	225-09490-01	Ion Gauge			X		
	261-00207-02	Gasket 14			X		
	225-20310-91	Pirani Gauge			X		
Main Unit	225-15718-91	DL Assy		X			
	036-19004-05	O-ring		X			
	228-48607-91	PEEK Frit		X			
	225-09340-11	Electron Multiplier			X		
	036-19004-53	Heater Flange O-ring			X		
	036-12523-00	Front Door O-ring			X		
	225-14985-01	Tune Solution		X			
ESI Probe	225-14948-91	ESI Capillary		X			
	228-33833-91	Peek Tubing		X			
	036-11205-00	O-ring		X			
APCI Probe	225-15877-92	Corona Needle		X			
	225-15845-91	APCI Capillary Pipe		X			
Miscellaneous	220-91239-00	LCMS Test Mix		X			

**Note:** These time frames are only suggestions and items may need to be replaced more frequently depending on sample throughput and use

Shop online ~ [store.shimadzu.com](http://store.shimadzu.com) or call Shimadzu at 800-477-1227

# LCMS-8030/8040

Triple Quadrupole Mass Spectrometer



LCMS-8030



LCMS-8040

## Inspection/Maintenance Schedule

	Part Number	Description	3 month Service	6 month Service	36 month Service	48 month Service	60 month Service
Vacuum System	220-97808-01	Pump Oil	X				
	220-91566-10	Oil Filter Element (MF30)		X			
	225-14179-01	Turbo Molecular Pump				X	
	225-09309-02	Rotary Pump, E2M28					X
	225-09490-01	Ion Gauge			X		
	261-00207-02	Gasket 14			X		
	225-20310-91	Pirani Gauge			X		
Main Unit	225-15718-91	DL Assy		X			
	036-19004-05	O-ring		X			
	228-48607-91	PEEK Frit		X			
	225-14168-01	Electron Multiplier			X		
	036-19004-53	Heater Flange O-ring			X		
	036-12527-00	Front Door O-ring			X		
	225-14122-01	Tune Solution		X			
ESI Probe	225-14948-91	ESI Capillary		X			
	228-33833-91	Peek Tubing		X			
	036-11205-00	O-ring		X			
APCI Probe	225-15877-92	Corona Needle		X			
	225-15845-91	APCI Capillary Pipe		X			
Miscellaneous	220-91239-03	8030 Test Standard		X			
	220-91239-00	LCMS Test Mix		X			

Note: These time frames are only suggestions and items may need to be replaced more frequently depending on sample throughput and use

# LCMS-8050

## Triple Quadrupole Mass Spectrometer



### Inspection/Maintenance Schedule

	Part Number	Description	3 month Service	6 month Service	36 month Service	48 month Service	60 month Service
Vacuum System	220-97808-01	Pump Oil	X				
	220-91566-10	Oil Filter Element (MF30)		X			
	225-14179-01	Turbo Molecular Pump				X	
	225-09309-02	Rotary Pump, E2M28					X
	225-09490-01	Ion Gauge			X		
	261-00207-02	Gasket 14			X		
	225-20310-91	Pirani Gauge			X		
Main Unit	225-15718-42	DL Assy 2		X			
	036-19004-05	O-ring		X			
	228-48607-91	PEEK Frit		X			
	225-14168-01	Electron Multiplier			X		
	036-19004-46	Heater Flange O-ring			X		
225-14122-01	Tune Solution		X				
ESI Probe	225-14948-91	ESI Capillary Assembly		X			
	228-33833-91	Peek Tubing		X			
	225-19560-41	ESI Heater Assy			X		
	036-11267-00	Front Door O-ring			X		
	036-11205-00	O-ring		X			
APCI Probe	225-15877-92	Corona Needle		X			
	036-11267-00	Front Door O-ring			X		
	225-15845-91	APCI Capillary Pipe		X			
Miscellaneous	220-91239-03	8030 Test Standard		X			
	220-91239-00	LCMS Test Mix		X			

Note: These time frames are only suggestions and items may need to be replaced more frequently depending on sample throughput and use

Shop online ~ [store.shimadzu.com](http://store.shimadzu.com) or call Shimadzu at 800-477-1227

# LCMS-8060

Triple Quadrupole Mass Spectrometer







## Inspection/Maintenance Schedule

	Part Number	Description	6 month Service	12 month Service	36 month Service	48 month Service	60 month Service
Vacuum System	220-91566-63	Leybold Pump Oil LVO700		X			
	220-91566-61	Exhaust Filter SV65-120B		X			
	220-91566-62	Maintenance Kit Leybold SV65			X		
	225-14179-01	Turbo Molecular Pump				X	
	220-91566-60	Leybold Rough Pump SV65					X
	225-09490-01	Ion Gauge			X		
	261-00207-02	Gasket 14			X		
	225-20310-91	Pirani Gauge			X		
Main Unit	225-15718-43	DL Assy 3	X				
	036-19004-05	O-ring	X				
	228-48607-91	PEEK Frit	X				
	225-14168-01	Electron Multiplier			X		
	036-19004-46	Heater Flange O-ring			X		
	225-14122-01	Tune Solution	X				
ESI Probe	225-14948-91	ESI Capillary Assembly	X				
	228-33833-91	Peek Tubing	X				
	225-19560-41	ESI Heater Assy			X		
	036-11267-00	Front Door O-ring			X		
	036-11205-00	O-ring	X				
APCI Probe	225-15877-92	Corona Needle	X				
	036-11267-00	Front Door O-ring			X		
	225-15845-91	APCI Capillary Pipe	X				
Miscellaneous	220-91239-03	8030 Test Standard	X				
	220-91239-00	LCMS Test Mix	X				

Note: These time frames are only suggestions and items may need to be replaced more frequently depending on sample throughput and use

## | Important

-  This publication may not be reproduced in whole or part without written permission from Shimadzu Corporation. Since Shimadzu products are frequently upgraded and improved, information in this publication is subject to change without notice.
-  Please read the instruction manual included with all products before use.
-  The replacement and maintenance periods listed in the guide are presented only as guidelines and not as a guarantee. Routine maintenance and inspection will vary depending on usage conditions.
-  Should you have questions on items listed in this guide, please contact your nearest Shimadzu representative.

## | Links

Shimadzu Scientific Instruments U.S. Webstore:  
<http://store.shimadzu.com>

Shimadzu Nexera Virtual Advisor:  
<https://nexerava.ssi.shimadzu.com/>

Shimadzu Prominence Virtual Advisor:  
<https://prominenceva.ssi.shimadzu.com/>



**SHIMADZU CORPORATION, International Marketing Division**

3. Kanda-Nishikicho 1-chome, Chiyoda-ku, Tokyo 101-8448, Japan

Phone: 81(3)3219-5641, Fax: 81(3)3219-5710

URL: [www.shimadzu.com](http://www.shimadzu.com)

**SHIMADZU SCIENTIFIC INSTRUMENTS**

7102 Riverwood Drive, Columbia, Maryland 21046, U.S.A.

Phone: 800-477-1227/410-381-1227, Fax: 410-381-1222

URL: [www.ssi.shimadzu.com](http://www.ssi.shimadzu.com)

Webstore URL: [store.shimadzu.com](http://store.shimadzu.com)

Company names, product/service names and logos used in this publication are trademarks and trade names of Shimadzu Corporation and its affiliates, whether or not they are used with trademark symbol "TM" or "®".

Third-party trademarks and trade names may be used in this publication to refer to either the entities or their products/services. Shimadzu disclaims any proprietary interest in trademarks and trade names other than its own.

For Research Use Only. Not for use in diagnostic procedures.

The contents of this publication are provided to you "as is" without warranty of any kind, and are subject to change without notice. Shimadzu does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication.