


StyleView® SV42 **Electronic Medical Records (EMR) Cart** with LCD Mount

StyleView powered carts provide electrical AC power for mobile point of care computing equipment in a healthcare environment. The carts are not intended to power medical products or devices. Outlets are provided to power information technology equipment only such as computer equipment and computer peripherals.

Features & Specifications	3
Set-up	4 - 14
Adjustment	12-13
Battery Charge/Discharge	14
Change Fuse (5 Amp) and Reset Circuit Breakers	15
Change Power System Batteries	16- 17
Cart Storage (long term/short term)	18
Ergonomics	19
Maintenance & Safety	19- 22
Dimensions	23



For service and warranty visit www.ergotron.com/warranty

For local customer care phone numbers visit: http://contact.ergotron.com

For the latest User Installation Guide and StyleLink Software Download please visit: www.ergotron.com




User's Guide - English Guía del usuario - Español Manuel de l'utilisateur - Français Gebruikersgids - Deutsch Benutzerhandbuch - Nederlands Guida per l'utente - Italiano Användarhandbok - svenska ユーザーガイド:日本語

用户指南 : 汉语









This Class A digital apparatus complies with Canadian ICES-003.

Cet appareil numérique de la classe A est conforme à la norme NMB-003 du Canada.

FCC Compliance Statement

The cart has been tested and found to comply with the limits for a Class A digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at his own expense.


Changes or modifications not expressly approved by Ergotron, Inc. could void the user's authority to operate the equipment.

 ${\it Please contact Ergotron for complete EMC compatibility information.}$



IMPORTANT! This product will need tension adjustments once installation is complete. Make sure all equipment is properly installed on the product before attempting range of motion or tension adjustments. Any time equipment is added or changed on this product resulting in a different mounted weight, you should repeat the adjustment steps to ensure safe and optimum operation. This product should move smoothly and easily through the full range of motion and stay where you set it. If movement is difficult or the product does not stay where you set it, follow the adjustment instructions to loosen or tighten the tension to create a smooth, easy motion. Depending on your product and the adjustment, it may take many turns to notice a difference.

Components




Tools Needed




2/23

Features & Specifications

Part Number	Power System
SV42-6301-1	Input: 120VAC/60 Hz, 5.1A; Output: 120VAC/60 Hz, 400VA, 300W. • The cart and power system are certified to UL 60601 and CAN/CSA-C22.2 601.1-M90
SV42-6301-2 SV42-6301-3 SV42-6301-B SV42-6301-C	Input: 230V~/50 Hz, 2.7A; Output: 230V~/50 Hz, 400VA, 300W. • The cart and power system are compliant with EN 60601-1.
SV42-6301-7 SV42-6301-8	Input: 220V~/60 Hz, 2.7A; Output: 220V~/60 Hz, 400VA, 300W.







- Height Adjustable LCD Mount attaches LCDs or tablet PC's with 75x75 or 100x100mm mounting interface
- 2. Worksurface 2a. Worksurface Lock and Release
- 3. User Interface
- 4. Secure Storage for Laptop, Thin Client or CPU
- 5. Front Handle
- 6. Height Adjustment Brake Handle
- 7. USB Hub connects keyboard and mouse USB cables
- 8. Keyboard tray slides out, tilts and allows for right or left mousing with attached mouse holder
- 9. Keyboard Light under Front Handle
- 10. Keyboard Light Switch
- 11. Cable Management and Storage for excess cables and power supplies
- 12. Storage Basket and Rear Handle
- 13. Front Locking Casters
- 14. Quick Reference Card
- 15. Power Cord Hooks
- **16.** Scanner Holder
- 17. Antimicrobial coating on worksurface and wrist rest
- 18. Power System

The StyleView AC Power System allows your power supply to travel with the cart. The Power System is integrated in the base of the cart and comes standard with 2 batteries, power module, User Interface (UI), outlet box and power cord.


- User Interface (UI): Allows power system output to be turned on or turned off, monitors battery charge remaining, and provides low battery charge audible alarm.
- Two 33 Ah Sealed Lead Acid, Absorbed Glass Mat, 12VDC batteries.
- The minimum operational temperature is 10° C (50° F) and the maximum operational temperature is 29° C (86° F). The recommended humidity range for operation is 5-95% rH.
- The recommended cart storage temperature is 15° C (59° F). At this temperature, the battery's age-related capacity loss is minimized. The minimum storage temperature is -20° C (-4° F) and the maximum storage temperature is 50° C (122° F). The recommended humidity range for storage is 5-95% rH.





888-24-175-G-01 rev. M • 02/16 3/23





CAUTION! Completely release brake engagement before raising or lowering the cart. Raising or lowering the cart with the brake partially engaged may cause product damage.













888-24-175-G-01 rev. M • 02/16 5/23











To increase space and improve airflow, power brick may be stored under the storage area.










Plug in power cables.

If you do not have power outlets located here, then follow steps 'a' - 'd' to route power cables.






888-24-175-G-01 rev. M • 02/16 7/23





Place computer and AC power adaptors in compartment. Do not place power bricks near




C

Plug the following factory connected cables into your computer.


USB: This cable runs from the USB Hub to your computer and uses your computer to power the USB Hub, Keyboard Lights and Fan. (NOTE: Your computer must be turned on for the USB Hub, Keyboard Lights and Fan to function).


 $\textbf{Optional StyleLink:} \ This \ USB \ cable \ runs \ from \ the \ power \ system \ to \ your \ computer \ for \ StyleLink \ Software. \ For \ more \ info \ on \ StyleLink \ visit: \ http://4support.ergotron.com\ .$




WARNING: Fan must always be running when computer is on. Operating computer without fan may lead to overheating, resulting in reduced equipment performance.

888-24-175-G-01 rev. M • 02/16 9/23











888-24-175-G-01 rev. M • 02/16 11/23 It is important that you adjust this product according to the weight of the mounted equipment as described in the following steps. Any time equipment is added or removed from this

product, resulting in a change in the weight of the mounted load, you should repeat these adjustment steps to ensure safe and optimum operation. Adjustments should move smoothly and easily through the full range of motion and stay where you set it. If adjustments are difficult and do not stay in the desired position, follow the instructions to loosen or tighten the tension to create a smooth, easy adjustment motion. Depending on your product and the adjustment, it may take several turns to notice a difference.










12/23 888-24-175-G-01 rev. M • 02/16



13/23 888-24-175-G-01 rev. M • 02/16

13

Initial Power on/Charge Battery (takes aproximately 7 hours to charge)

a P


Plug Cart's Power Cord into wall outlet.



Do Not stretch coiled cord further than 8 feet (2.5 meters), damage to the cord may occur.



CAUTION: There is no on/off switch on this equipment; the AC power cord is the only power disconnect. The socket outlet should be easily accessible and should be installed near the equipment.




With cart's power cord plugged into the wall outlet, wait until cart is at 100% charge. (takes aproximately up to 7 hours to charge)



This power system interface will alert you to the percentage of charge remaining in the cart battery with a series of steady or flashing red, yellow or green lights, and an alarm that will beep when charge gets below 30%. Remember, the battery needs to be charged to 100% every day, and you can use the cart while charging, so plug-in cord as often as possible to avoid running out of power!

NOTE: Put monitor in power save mode to optimize battery run time.




14

Keyboard Light


1. Turn Computer on.


2. Test Keyboard Light.

Keyboard Light will automatically turn off after 15 minutes if not manually turned off.




14/23











Before reseting circuit breakers, contact Ergotron to determine cause of trip.





888-24-175-G-01 rev. M • 02/16 15/23

WARNING: RISK OF ELECTRICAL DISCHARGE

- Do NOT swap battery cables, doing so will cause arcing and trip the circuit breaker.
- Do not remove or replace the batteries while cart is located in an oxygen rich or hazardous environment, arcing may occur and cause combustion.
- Replace first battery before removing second battery to reduce the risk of cables touching terminals and causing arcing.
- · Always replace both batteries!
- Replace first battery before removing second battery.
- Replace batteries with same Amp/Hour rating batteries only.
- Only Ergotron-specified batteries may be used in the StyleView Power System. Please call customer care for more details.
- Recycle battery or contact Ergotron for proper battery disposal auidelines.



33 Amp-Hour Batteries



WARNING AVERTISSEMENT

BATTERY REPLACEMENT

The power module is configured for two, 33 -Ah batteries.
Only the following batteries are compatible with this

- B&B EP33-12
- SBS S-12330 • Vision 6FM33D
- · Longway 6FM33G Vision 6FM33HD-X
- PBQ PBQ33-12L Werker WKA12-33C Werker WKA12-35C
- Duracell Dura 12-35C

Installing batteries other than the 33 -Ah listed above will void the product warranty. Failure to heed this warning may result in severe damage to batteries, power module and possible fire hazard.

REMPLACEMENT DE LA BATTERIE

Le module d'alimentation est configuré pour deux batteries 33 A-h.
Seules les batteries suivantes sont compatibles avec ce système :

- B&B EP33-12
- SBS S-12330
- Longway 6FM33G Vision 6FM33D Vision 6FM33HD-X
- PRO PRO33-121 • Werker WKA12-33C • Werker WKA12-35C

• Duracell Dura 12-35C


L'installation de batteries autres que les batteries 33 A-h citées ci-dessus annulera la garantie du produit. Si vous ne vous conformez pas à cette précaution d'utilisation, les batteries et le module d'alimentation risquent d'être sérieusement endommagés, ce qui pourrait provoquer un incendie.



www.ergotron.com

822-386








Disconnect Power System from power source.




Turn power system off by holding down the AC Outlet Power button for 1 - 3 seconds. Power light will shut off.






Turn both circuit breakers to "OFF".




















Recycle battery or contact **Ergotron for proper battery** disposal guidelines.










16/23




Follow Battery Charge/Discharge Initial Power on steps a, b, c.


888-24-175-G-01 rev. M • 02/16 17/23


Short Term Storage - If the Power System will be idle for up to 3 months, the battery should be fully charged before storage.

Long Term Storage - If the Power System will be idle for 3 months or more, the battery should be removed from the cart and recharged during storage.



Long Term Storage - If the Power System will be idle for 3 months or more, the battery should be removed from the cart and recharged during storage.





18/23 888-24-175-G-01 rev. M • 02/16

Ergonomics


Working

customize - to your size

- 1 Set top of monitor screen about one inch below eye level Release brake and lift or lower riser as needed.
- 2 Tilt screen for comfortable viewing and to reduce eye and neck strain.
- 3 Pull keyboard tray forward and position mouse tray on right or left, as needed.
- 4 Work with elbows bent at about 90° to minimize muscle strain.
- 5 If the riser moves up and down with difficulty, or if it drifts out of set position, consult the product manual for adjustment information.

6 Stay in charge!

Powered carts should be plugged into outlet as often as possible to keep battery charged and computer running.




Moving

stow - before you go

- 1 During normal movement, release brake and lower worksurface to lowest position for optimal stability and unobstructed view.
- 2 Tuck away open trays and return mouse, scanner and other accessories to their places.
- 3 Unlock both front casters.
- 4 Push cart from rear with elbows bent at about 90° to maximize control and minimize muscle strain.
- 5 Don't run out!

Before moving, make sure cord is unplugged from outlet and hooked to basket for safe travel. Remember, charge battery fully 100% every day!



Maintenance & Safety Hazard Symbols Review

The Meaning of Symbols appearing in this Guide, on the Cart or on the Power System

These symbols alert you to a safety condition that demands your attention. You should be able to recognize and understand the significance of the following Safety Hazards if you encounter them on the Cart or within Cart documentation such as this Set-up Guide.

Symbol	Signal Word/ Color	Level of Hazard	
	DANGER	Indicates an imminently hazardous situation which, if not avoided, will result in death or serious injury.	
	WARNING	Indicates a potentially hazardous situation which, if not avoided, could result in death or serious injury.	
	CAUTION	Indicates a potentially hazardous situation which, if not avoided, may result in minor or moderate injury.	
	CAUTION	Used without the safety alert symbol indicates a potentially hazardous situation which, if not avoided, may result in property damage.	
	INSTRUCTIONS	Follow operating instructions.	
$\bigcirc i$	INSTRUCTIONS	Follow operating instructions.	
(1)	POWER	"ON" / "OFF" (push-push) NOTE: Each position "ON" / "OFF" is a stable position.	



EQUIPMENT & ACCESSORIES DISPOSAL

- 1. Please dispose of all batteries in accordance with local law
- 2. All Electronics should be recycled through an electronics recycler.
- 3. Remaining plastics and metals can be recycled through a commercial recycler.

888-24-175-G-01 rev. M • 02/16 19/23

Maintenance & Safety



Cleaning and Maintenance

The following procedures are not guaranteed to control infection. The hospital infection control administrator or epidemiologist should be consulted regarding cleaning procedures and processes.

🔼 To avoid risk of electric shock, do not expose electrical components to water, cleaning solutions or other potentially corrosive liquids or substances.

Do not immerse Cart or Cart components in liquid or allow liquids to flow into the Cart. Wipe all cleaners off surface immediately using a damp cloth. Thoroughly dry surface after cleaning

Do not use flammable cleaners on Cart surfaces due to close proximity of electrical power and equipment.

All paints and plastic Cart components will withstand cleaning by most commonly used, diluted, non-abrasive solutions such as quaternary ammonia compounds, ammonia enzyme cleaners, bleach or alcohol solutions.

- Pen and permanent and dry erase markers can be removed with 91% isopropyl alcohol and a soft cloth.
- · lodine stains can be removed with commonly used cleaners and a soft cloth.
- Never use steel wool or other abrasive materials that will damage the surface finish.

Do not use strong solvents such as trichloroethylene and acetone. These solvents will damage the surface finish.
It is recommended that any cleaning solution be tested on a small, inconspicuous area to ensure surface is not harmed.

Adjustment, Service, Replacement - DO NOT attempt to adjust, service or replace any part of the StyleView Cart unless directed to do so through Ergotron-approved documentation (i.e. installation instructions). Only Ergotron, Inc. or an Ergotron-certified entity may adjust, service or replace StyleView Cart components. If any component on the Cart is missing or damaged, the Cart must not be used, contact Ergotron Customer Care immediately to request a replacement part.

Cables - Keep cables neatly organized on the Cart (a variety of solutions are provided with your cart for this purpose). Excess cables should be routed away from moving components with cable clips. Review Cable Routing section of this guide, or contact Ergotron Customer Care for more information.

Casters - Check casters periodically to make sure they are clean and free of debris that would prevent smooth travel. Avoid moving Cart across uneven, dirty or damaged surfaces.

Customer Equipment- Make sure equipment is balanced and mounted securely to Cart. Do not reposition Cart components on riser or tower unless instructed to do so in the installation instructions. Moving Cart components too high or too low on the Riser may create an unstable condition, leading to equipment damage or even personal injury. Contact Ergotron Customer Care for information about moving Cart components.

Safety Alerts Associated with this Product

The following Warnings/Cautions appear in this reference guide or on the cart:

NOTE: Failure to adhere to these guidelines may result in equipment damage or personal injury.

CAUTION: The lift brake helps stablilize the worksurface and keyboard tray during normal use but it DOES NOT increase load capacity. DO NOT load riser with equipment totaling more than the maximum weight capacity specified by Ergotron. Ensure optimum lift function by testing and if necessary, re-adjusting tension whenever the weight mounted to the riser changes (i.e., equipment is removed or added). See "Set Riser Lift Tension" adjustment instructions.

CAUTION: Do not operate StyleView Cart with missing or damaged components! Do not remove, modify or substitute Cart components without consulting Ergotron. If you encounter problems with Cart installation or operation, contact Ergotron Customer Care.

CAUTION: DO NOT overtighten fasteners. Overtightening may cause damage to your equipment.

WARNING: Stored Energy Hazard: The worksurface lift mechanism is under tension and will move up rapidly, on its own, as soon as attached equipment is removed. For this reason, DO NOT remove equipment unless the worksurface has been moved to the highest position on the tower! Failure to follow this instruction may result in serious personal injury and/or equipment damage! When Shipping the cart, set the worksurface lift mechanism to the lowest tension setting.

CAUTION: DO NOT loosen, tighten or remove any other nuts or bolts on the riser or top of tower. Tampering with nuts or bolts may result in an unstable Cart, leading to equipment damage and/or personal injury.

CAUTION: Release Lift Brake before moving work surface! Moving work surface while Lift Brake is engaged may cause serious damage to Lift Engine.

MARNING: In the event that repair of the StyleView Cart is needed, contact Ergotron Customer Care immediately. Cart repair can only be performed by Ergotron, Inc. or by an Ergotron authorized agent.

WARNING: This cart is not intended for use in a flammable, anesthetic mixture or oxygen rich environment.

Configuration & Safety

Additional multiple socket-outlet or extension cord shall not be connected to the medical system.

When used in a Medical Electrical system, connect only equipment that complies with IEC, ISO, UL/ANSI, or CSA standards that are relevant to that equipment.

👫 Risk of shock or personal injury when connecting non-medical equipment supplied as part of a system directly to the wall outlet when non-medical equipment is intended to be supplied by the multiple socket outlet.

🖺 Risk of shock or personal injury when connecting any equipment that has not been supplied as part of the medical system to the multiple socket outlet.

20/23 888-24-175-G-01 rev. M • 02/16

Recommended Periodic Inspection and Maintenance			
Component	Action	How often	By whom
UI, Ethernet, USB cables	Inspect for wear, pinching, bad connectors	Monthly	Any user
Fan on side of CPU compartment	Inspect for dust at intake, vacuum as required using a vacuum cleaner that DOES NOT generate ESD (Electrostatic Discharge)	Monthly	Any user
Casters	Inspect for wear and debris	Monthly	Any user
Maximum Load	Inspect to ensure that maximum recommended loads are not exceeded	Daily	Any user
Power System SLA batteries	Replace lead acid batteries ^{1,2}	When instructed by StyleLink or when battery runtime is 80% of original runtime	IT Personnel
Power Module	Inspect for dust at intake, vacuum as required using a vacuum cleaner that DOES NOT generate ESD (Electrostatic Discharge)	Monthly	IT Personnel
Battery harness	Inspect for wear, cracking, pinching, or other damage	Monthly	IT Personnel
Coiled Cord	Inspect for wear, damage, or stretching.	Weekly	Any user

- 1. Please dispose of all batteries in accordance with local law
- 2. Always replace with similar size battery (33 A-h to replace 33 A-h) and always replace in pairs

Earth Bond Test: To ensure safety grounding between the power cord ground connection and any accessible metal parts on the cart, although not mandatory, the following test is recommended to be performed every two years using a calibrated medical device safety analyzer. The procedure is as

- 1. Disconnect the cart from mains power (unplug cart from the
- 2. Remove power from internal power system outlets by pressing the power button on the Power System Interface.
- 3. Set up Earth Bond Test per the medical device safety analyzer instructions.
- 4. Connect the Cart power cord to the medical device safety analyzer.
- 5. Remove cover to access Lift Tension Adjustment point.
- 6. Attach second medical safety device analyzer probe to Lift Tension Adjustment point on Cart.
- 7. Perform Test (25 amps).
- 8. Ensure Earth Bond is less than or equal to 0.2-ohms.
- 9. Remove medical device safety analyzer connections to Cart, replace cover over Lift Tension Adjustment point, return the Cart to service.

Insulation Resistance Test: To ensure mains power lines are adequately insulated from earth ground, the following test, although not mandatory, is recommended to be performed annually using a calibrated medical device safety analyzer. The procedure is as follows:

- 1. Disconnect the cart from mains power (unplug cart from the
- 2. Remove power from internal power system outlets by pressing the power button on the Power System Interface.
- 3. Set up the Insulation Resistance Test per the medical device safety analyzer instructions.
- 4. Connect the Cart power cord to the medical device safety analyzer.
- 5. Perform test (500 V).
- 6. Ensure Insulation Resistance is greater than or equal to 1
- 7. Remove medical device safety analyzer connections to Cart and return Cart to service.



- Inspect Cord Thoroughly Before Each Use. DO NOT USE IF DAMAGED.
 Do Not Plug More Than SPECIFIED NUMBER OF WATTS Into Power System.
 Do Not Run Cord Through Doorways, Holes in Ceilings, Walls or Floors.
 FULLY INSERT Plug Into Outlet.
 Do Not Remove, Bend or Modify Any Metal Prongs or Pins of Cord.
 Do Not Use Excessive Force to Make Connections.
 Keep Away From Water. DO NOT USE WHEN WET.

- Reep Away From Water. LO NOT USE WHEN WE I.

 Keep Children Away From Cord.

 Do Not Plug Cord into An Extension Cord.

 AVOID OVERHEATING. Uncoil Cord and Do Not Cover It With Any Material.

 Do Not Drive, Drag or Place Objects Over Cord. Do Not Walk On Cord.

 GRASP PLUG to REMOVE FROM Outlet. Do Not Unplug By Pulling On Cord.

888-24-175-G-01 rev. M • 02/16 21/23 Ergotron Electromagnetic Guidance and Manufacturer's Declaration

Guidance and Manufacturer's Declaration – Electromagnetic Emissions

The Powered Computer Cart is intended for use in the electromagnetic environment specified below. The customer or the user of the Powered Computer Cart should assure that it is used in such an environment.


Emissions Test	Compliance	Electromagnetic environment – guidance	
RF Emissions CISPR 11	Group 1	The Powered Computer Cart uses RF energy only for its internal function. Therefore, its RF emissions are very low and unlikely to cause any interference in nearby electronic equipment.	
RF Emissions CISPR 11	Class A	The Powered Computer Cart is suitable for use in all establishments other than domestic and those directly	
Harmonic Emissions IEC 61000-3-2	Class A	connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.	
Voltage fluctuations/ flicker emissions IEC 61000-3-3	Complies	, puiposes.	

Guidance and Manufacturer's Declaration – Electromagnetic Immunity

The Powered Computer Cart is intended for use in the electromagnetic environment specified below. The customer or the user of the Powered Computer Cart should assure that it is used in such

Immunity Test	IEC 60601 Test Level	Compliance Level	Electromagnetic environment - guidance
Electrostatic Discharge (ESD) IEC 61000-4-2	±6 kV contact ±8 kV air	Complies	Floors should be wood, concrete, or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%
Electrical Fast Transient/Burst IEC 61000-4-3	±2 kV for power supply lines ±1 kV for input/output lines	Complies	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV differential mode ±2 kV common mode	Complies	Mains power quality should be that of a typical commercial or hospital environment
Voltage Dips, short interruptions, and voltage variations on power supply input lines IEC 61000-4-11	$ \begin{array}{l} <5\% \ U_T \\ (>95\% \ dip \ in \ U_T) \ for \ 0.5 \ cycle \\ 40\% \ U_T \\ (60\% \ dip \ in \ U_T) \ for \ 5 \ cycles \\ 70\% \ U_T \\ (30\% \ dip \ in \ U_T) \ for \ 25 \ cycles \\ <5\% \ U_T \\ (>95\% \ dip \ in \ U_T) \ for \ 5 \\ seconds \end{array} $	Complies	Mains power quality should be that of a typical commercial or hospital environment.
Power Frequency (50/60 Hz) Magnetic Field IEC 61000-4-8	3 A/m	Complies	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
Note: U _T is the AC mains voltage prior to application of the test level			

22/23 888-24-175-G-01 rev. M • 02/16



ERGOTRON®

While Ergotron, Inc.makes every effort to provide accurate and complete information on the installation and use of its products, it will not be held liable for any editorial errors or omissions (including those made in the process of translation from English to another language), or for incidental, special or consequential damages of any nature resulting from furnishing this instruction and performance of equipment in connection with this instruction. Ergotron, Inc. reserves the right to make changes in the product design and/or product documentation without notification to its users. For the most current product information, or to know if this document is available in languages other than those herein, please contact Ergotron. No part of this publication may be reproduced, stored in a retrieval system, or transmitted in any form or by any means, electronic, mechanical photocopying, recording or otherwise without the prior written consent of Ergotron, Inc., 1181 Trapp Road, Eagan, Minnesota, 55121, USA Patents Pending and Patented U.S. & Foreign. Ergotron is a registered trademark of Ergotron, Inc.

© 2016 Ergotron, Inc. All rights reserved.

Americas Sales and Corporate Headquarters

1181 Trapp Rd. St. Paul, MN USA (800) 888-8458 +1-651-681-7600 www.ergotron.com

sales@ergotron.com **APAC Sales**

Tokyo, Japan www.ergotron.com

EMEA Sales

Amersfoort, The Netherlands +31 33 45 45 600 www.ergotron.com

info.eu@ergotron.com

Worldwide 0EM Sales

www.ergotron.com info.oem@ergotron.com

apaccustomerservice@ergotron.com

888-24-175-G-01 rev. M • 02/16 23/23