



Alert number

**A12/01421/21**

Product

**Particle filter mask**



Risk type **Health risk / other**

**The particle/filter retention of the material is insufficient (measured values: as low as 83%). Consequently, the product does not fulfil the health and safety requirements, and an excessive amount of particles or microorganisms might pass through the mask, increasing the risk of infection; thus, even if combined with other recommended measures, it may not properly protect.**

**The product does not comply with the Personal Protective Equipment (PPE) Regulation and the relevant European standard EN 149.**

Category **Protective equipment**

Type **Consumer**

Description **White protective respiratory filtration half-face mask with the brand name, the EN standard, FFP2 and the NR marking printed on.**

Packaging description **Box with logo, description of the product, quantity (50 units) and indication of the EN149:2001 + A1:2009 and the filter efficiency (>94%). Instructions for use in English.**

Brand **Biosis Healing**

Name **Biosis Healing Protective Mask**

Type / number of model **BH-YEG-15.5x10.5**

Batch number **620071013**

Weekly report number	<b>Report-2021-41</b>
Alert submitted by	<b>Ireland</b>
Is the product counterfeit?	<b>Unknown</b>
Country of origin	<b>People's Republic of China</b>
Measures ordered by public authorities	<b>Destruction of the product Other</b>
Date of entry into force	<b>05/08/2021</b>