

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PENNSAUKEN TWSP. P.D. Serial No.: ARUM-0066
Calibration File No.: 02650 Calib. Date: 08/26/2019 Calib. No.: 00036
Certification File No.: 02651 Cert. Date: 08/26/2019 Cert. No.: 00032
Linearity File No.: 02652 Lin. Date: 08/26/2019 Lin. No.: 00032
Solution File No.: 02696 Soln. Date: 11/25/2019 Soln. No.: 00283
Sequential File No.: 02696 File Date: 11/25/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 0201

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:23S		
Control 1 EC	0.101%	15:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:24S		
Control 2 EC	0.099%	15:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:26S		
Control 3 EC	0.099%	15:26S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: PETER

MI: E

Signature: _____



Badge No.: 187

Date: 11/25/2019

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02650 Calib. Date: 08/26/2019 Calib. No.: 00036
Certification File No.: 02651 Cert. Date: 08/26/2019 Cert. No.: 00032
Linearity File No.: 02652 Lin. Date: 08/26/2019 Lin. No.: 00032
Solution File No.: 02695 Soln. Date: 11/08/2019 Soln. No.: 00282
Sequential File No.: 02695 File Date: 11/08/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 0255

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:13S		
Control 1 EC	0.100%	16:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:14S		
Control 2 EC	0.099%	16:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:15S		
Control 3 EC	0.099%	16:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: LT M Henkel 149

Badge No.: 149

Date: 11/08/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02650	Calib. Date:	08/26/2019
Certification File No.:	02651	Cert. Date:	08/26/2019
Linearity File No.:	02652	Lin. Date:	08/26/2019
Solution File No.:	02682	Soln. Date:	10/11/2019
Sequential File No.:	02682	File Date:	10/11/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUN S3-0341
Control Solution %:	0.100%			Expires:	06/10/2021
Solution Control Lot:	19150			Bottle No.:	0254

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:44D		
Control 1 EC	0.098%	02:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:45D		
Control 2 EC	0.097%	02:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:46D		
Control 3 EC	0.099%	02:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____



#219

Badge No.: 219

Date: 10/11/2019

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02650 Calib. Date: 08/26/2019 Calib. No.: 00036
Certification File No.: 02651 Cert. Date: 08/26/2019 Cert. No.: 00032
Linearity File No.: 02652 Lin. Date: 08/26/2019 Lin. No.: 00032
Solution File No.: 02675 Soln. Date: 10/05/2019 Soln. No.: 00280
Sequential File No.: 02675 File Date: 10/05/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 0304

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:15D		
Control 1 EC	0.099%	00:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:17D		
Control 2 EC	0.097%	00:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:18D		
Control 3 EC	0.099%	00:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

PH. A. Mireles #219

Badge No.: 219

Date: 10/05/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02650	Calib. Date: 08/26/2019
Certification File No.:	02651	Calib. No.: 00036
Linearity File No.:	02652	Cert. Date: 08/26/2019
Solution File No.:	02665	Cert. No.: 00032
Sequential File No.:	02665	Lin. Date: 08/26/2019
		Lin. No.: 00032
		Soln. Date: 09/14/2019
		Soln. No.: 00279
		File Date: 09/14/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0258

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:43D		
Control 1 EC	0.098%	23:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:45D		
Control 2 EC	0.098%	23:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	23:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:46D		
Control 3 EC	0.097%	23:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

M. Mireles #219

Badge No.: 219

Date: 09/14/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02568	Calib. Date: 04/11/2019
Certification File No.:	02569	Calib. No.: 00035
Linearity File No.:	02570	Cert. Date: 04/11/2019
Solution File No.:	02645	Cert. No.: 00031
Sequential File No.:	02645	Lin. Date: 04/11/2019
		Lin. No.: 00031
		Soln. Date: 08/24/2019
		Soln. No.: 00277
		File Date: 08/24/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 1490

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:17D		
Control 1 EC	0.100%	22:17D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	22:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:18D		
Control 2 EC	0.098%	22:19D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:19D		
Control 3 EC	0.099%	22:20D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRISTOW

First Name: MARK

MI: V

Signature: _____

PTL. Bristow #227

Badge No.: 227

Date: 08/24/2019

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02568 Calib. Date: 04/11/2019 Calib. No.: 00035
Certification File No.: 02569 Cert. Date: 04/11/2019 Cert. No.: 00031
Linearity File No.: 02570 Lin. Date: 04/11/2019 Lin. No.: 00031
Solution File No.: 02632 Soln. Date: 08/10/2019 Soln. No.: 00276
Sequential File No.: 02632 File Date: 08/10/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 1158

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:52D		
Control 1 EC	0.098%	06:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:54D		
Control 2 EC	0.097%	06:54D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	06:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:55D		
Control 3 EC	0.097%	06:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: M. Mireles 219

Badge No.: 219

Date: 08/10/2019

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02568	Calib. Date: 04/11/2019
Certification File No.:	02569	Calib. No.: 00035
Linearity File No.:	02570	Cert. Date: 04/11/2019
Solution File No.:	02619	Cert. No.: 00031
Sequential File No.:	02619	Lin. Date: 04/11/2019
		Lin. No.: 00031
		Soln. Date: 08/02/2019
		Soln. No.: 00275
		File Date: 08/02/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 0936

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:15D		
Control 1 EC	0.098%	10:16D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		
Control 2 EC	0.097%	10:17D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		
Control 3 EC	0.097%	10:18D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL	First Name: GERALD	MI: F.
Signature: <u></u>	149	Badge No.: 149
		Date: 08/02/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02568	Calib. Date: 04/11/2019
Certification File No.:	02569	Calib. No.: 00035
Linearity File No.:	02570	Cert. Date: 04/11/2019
Solution File No.:	02611	Cert. No.: 00031
Sequential File No.:	02611	Lin. Date: 04/11/2019
		Lin. No.: 00031
		Soln. Date: 07/19/2019
		Soln. No.: 00274
		File Date: 07/19/2019

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0340
Control Solution %:	0.100%		Expires: 09/17/2020
Solution Control Lot:	18290		Bottle No.: 1052

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:33D		
Control 1 EC	0.101%	15:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:34D		
Control 2 EC	0.100%	15:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:35D		
Control 3 EC	0.100%	15:36D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBONE

First Name: BRIAN

MI: M

Signature: _____

T.P.I. Gambone # 7029

Badge No.: 7029

Date: 07/19/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02568	Calib. Date:	04/11/2019
Certification File No.:	02569	Cert. No.:	00035
Linearity File No.:	02570	Cert. Date:	04/11/2019
Solution File No.:	02609	Lin. No.:	00031
Sequential File No.:	02609	Lin. Date:	04/11/2019
		Soln. No.:	00273
		Soln. Date:	06/09/2019
		File Date:	06/09/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0340
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	1060

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:42D		
Control 1 EC	0.097%	09:43D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:44D		
Control 2 EC	0.098%	09:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45D		
Control 3 EC	0.098%	09:46D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERN

First Name: RAYMOND

MI: T

Signature: _____

Ray Kern #181

Badge No.: 181

Date: 06/09/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02568	Calib. Date: 04/11/2019
Certification File No.:	02569	Calib. No.: 00035
Linearity File No.:	02570	Cert. Date: 04/11/2019
Solution File No.:	02597	Cert. No.: 00031
Sequential File No.:	02597	Lin. Date: 04/11/2019
		Lin. No.: 00031
		Soln. Date: 05/28/2019
		Soln. No.: 00272
		File Date: 05/28/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1040

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:19D		
Control 1 EC	0.101%	10:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20D		
Control 2 EC	0.100%	10:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22D		
Control 3 EC	0.099%	10:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKBL

First Name: GERALD

MI: F

Signature: _____

Handwritten signature: G. Henkel 149

Badge No.: 149

Date: 05/28/2019



Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02568	Calib. Date: 04/11/2019
Certification File No.:	02569	Calib. No.: 00035
Linearity File No.:	02570	Cert. Date: 04/11/2019
Solution File No.:	02596	Cert. No.: 00031
Sequential File No.:	02596	Lin. Date: 04/11/2019
		Lin. No.: 00031
		Soln. Date: 05/25/2019
		Soln. No.: 00271
		File Date: 05/25/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1166

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:31D		
Control 1 EC	-.---%	07:32D	22.7°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	07:32D	22.7°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERN	First Name: RAYMOND	MI: T
Signature: <u></u>		Badge No.: 181
		Date: 05/25/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02568	Calib. Date:	04/11/2019
Certification File No.:	02569	Cert. Date:	04/11/2019
Linearity File No.:	02570	Lin. Date:	04/11/2019
Solution File No.:	02595	Soln. Date:	05/25/2019
Sequential File No.:	02595	File Date:	05/25/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0340
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	1165

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:59D		
Control 1 EC	-.---%	04:00D	23.4°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	04:00D	23.4°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: I

Signature: _____



219

Badge No.: 219

Date: 05/25/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02568	Calib. Date: 04/11/2019
Certification File No.:	02569	Calib. No.: 00035
Linearity File No.:	02570	Cert. Date: 04/11/2019
Solution File No.:	02583	Cert. No.: 00031
Sequential File No.:	02583	Lin. Date: 04/11/2019
		Lin. No.: 00031
		Soln. Date: 05/04/2019
		Soln. No.: 00269
		File Date: 05/04/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1277

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:24D		
Control 1 EC	0.098%	16:25D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:26D		
Control 2 EC	0.098%	16:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:27D		
Control 3 EC	0.098%	16:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 05/04/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02469	Calib. Date: 10/24/2018
Certification File No.:	02470	Calib. No.: 00034
Linearity File No.:	02471	Cert. Date: 10/24/2018
Solution File No.:	02563	Cert. No.: 00030
Sequential File No.:	02563	Lin. Date: 10/24/2018
		Lin. No.: 00030
		Soln. Date: 03/30/2019
		Soln. No.: 00267
		File Date: 03/30/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1288

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:29S		
Control 1 EC	0.098%	20:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:31S		
Control 2 EC	0.097%	20:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:32S		
Control 3 EC	0.097%	20:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRISTOW

First Name: MARK

MI: V

Signature: _____

PTL. Bristow #227

Badge No.: 227

Date: 03/30/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02469	Calib. Date: 10/24/2018
Certification File No.:	02470	Calib. No.: 00034
Linearity File No.:	02471	Cert. Date: 10/24/2018
Solution File No.:	02547	Cert. No.: 00030
Sequential File No.:	02547	Lin. Date: 10/24/2018
		Lin. No.: 00030
		Soln. Date: 03/01/2019
		Soln. No.: 00266
		File Date: 03/01/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1055

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:21S		
Control 1 EC	0.100%	02:21S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:22S		
Control 2 EC	0.099%	02:23S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:23S		
Control 3 EC	0.099%	02:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRISTOW

First Name: MARK

MI: V

Signature: _____

PTL. Bristow #227

Badge No.: 227

Date: 03/01/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02469	Calib. Date: 10/24/2018
Certification File No.:	02470	Calib. No.: 00034
Linearity File No.:	02471	Cert. Date: 10/24/2018
Solution File No.:	02535	Cert. No.: 00030
Sequential File No.:	02535	Lin. Date: 10/24/2018
		Lin. No.: 00030
		Soln. Date: 02/10/2019
		Soln. No.: 00265
		File Date: 02/10/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0662

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:06S		
Control 1 EC	0.098%	11:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07S		
Control 2 EC	0.098%	11:08S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09S		
Control 3 EC	0.099%	11:09S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: CIACCIA

First Name: MICHAEL

MI: -

Signature: _____



#172

Badge No.: 172

Date: 02/10/2019



Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02469 Calib. Date: 10/24/2018 Calib. No.: 00034
Certification File No.: 02470 Cert. Date: 10/24/2018 Cert. No.: 00030
Linearity File No.: 02471 Lin. Date: 10/24/2018 Lin. No.: 00030
Solution File No.: 02524 Soln. Date: 01/25/2019 Soln. No.: 00264
Sequential File No.: 02524 File Date: 01/25/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 03/13/2020
Solution Control Lot: 18090 Bottle No.: 0092

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:40S		
Control 1 EC	0.100%	23:40S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:41S		
Control 2 EC	0.099%	23:42S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:42S		
Control 3 EC	0.099%	23:43S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: TIMOTHY

MI: I

Signature: _____

Timothy Mireles 219

Badge No.: 219

Date: 01/25/2019

