

**CERTIFICATE OF ANALYSIS**

PRODUCT SR0244E
OCLA (ISO)DIFFERENTIAL SUPPLEMENT

LOT NUMBER 2362170

EXPIRY DATE 2019.04.30

DATE OF MANUFACTURE 2018.07.04

Delivery/Customer information

Date Printed
2018.07.10
Delivery No.

Customer
Customer Order Number

| Physical Characteristics | Results | Specification |
|--------------------------|----------------------------------|----------------------------------|
| Appearance | Bi-phasic or opaque/cream liquid | Bi-phasic or opaque/cream liquid |

| Microbiological Performance | Control cfu | Test cfu | Recovery % | Test Description |
|-----------------------------|-------------|----------|------------|------------------|
|-----------------------------|-------------|----------|------------|------------------|

Testing has been performed in accordance with ISO11133:2014

Tested in Chromogenic Listeria Agar (ISO) CM1084 and OCLA (ISO) Selective Supplement SR0226

Aerobic incubation at 37 ± 2°C for 44 ± 4 hours

| | | | | |
|--|---------|----|-----|------------------------|
| <i>L. monocytogenes</i> ATCC@13932 WDCM00021 | 71 | 67 | 94 | Blue cols with halo |
| <i>L. monocytogenes</i> ATCC@35152 WDCM00109 | 80 | 86 | 108 | Blue cols with halo |
| <i>L. innocua</i> ATCC@33090 WDCM00017 | 1.8E+03 | - | - | Blue cols without halo |
| <i>E. coli</i> ATCC@25922 WDCM00013 | >1E+04 | 0 | - | No growth |
| <i>E. coli</i> ATCC@8739 WDCM00012 | >1E+04 | 0 | - | No growth |
| <i>E. faecalis</i> ATCC@29212 WDCM00087 | >1E+04 | 0 | - | No growth |
| <i>E. faecalis</i> ATCC@19433 WDCM00009 | >1E+04 | 0 | - | No growth |

Control Media: Tryptone Soya Agar or Columbia Blood Agar Base enriched with 5% v/v horse blood, where appropriate

A satisfactory result is represented by recovery of positive strains equal to or greater than 50% of the control medium.
Refer to product specification for full details.



Tested by the Quality Control Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096



CERTIFICATE OF ANALYSIS

PRODUCT SR0244E
LOT NUMBER 2362170
OCLA (ISO)DIFFERENTIAL SUPPLEMENT

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release.
Lot Accepted. 2018.07.10

Shannon Bearpark
Site Quality Manager

ATCC is a registered trade mark of the American Type Culture Collection.
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory
OXOID LIMITED
Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com